

BC-RET

HEMATOLOGY CONTROLS

CONTROL

ASSAY VALUES AND EXPECTED RANGES

LOT ME0924A



2024-11-10

Instrument	Parameter	Low		Normal		High	
		LOT	ME0924A-1	LOT	ME0924A-2	LOT	ME0924A-3
element HT5+	Ret# x10 ³ /μL	67.0	± 64.4	216.7	± 99.0	188.1	± 99.0
QC Mode	Ret%	1.12	± 1.12	4.62	± 2.00	8.25	± 3.50
	IRF %	13.5	± 15.0	13.2	± 13.0	10.1	± 25.0
	LFR %	86.5	± 15.0	86.8	± 15.0	89.9	± 25.0
	MFR %	12.5	± 12.0	12.1	± 10.0	7.9	± 22.0
	HFR %	1.0	± 10.0	1.1	± 10.0	2.2	± 10.0
	RHE pg	27.1	± 12.0	25.4	± 12.0	20.8	± 12.0
	IPF %	7.7	± 15.0	7.6	± 15.0	7.6	± 15.0

Before using, refer to the instruction sheet for mixing directions.